

☐ SECRET

☒ CONFIDENTIAL

☐ UNCLASSIFIED

Approved For Release 2005/05/02 : CIA-RDP78B04770A001200010092-8
CONTRACT INSPECTION REPORT

TASK NO.

25X1

TO:

DATE

1 October 1968

INSPECTION REPORT NO. (If final, so state)

1

ESTIMATED COMPLETION DATE

1 August 1969

NAME OF CONTRACTOR

Prototype Twin Stage On-Line PI Comparator and a Precise Measuring Comparator

THE CONTRACTOR IS ON SCHEDULE

☐

YES

☒

NO

PER CENT OF WORK COMPLETED - 5%

PER CENT OF FUNDS EXPENDED - 3.2% (24 Aug 1968)

THE CONTRACTOR WILL PROBABLY REMAIN WITHIN ALLOCATED FUNDS ☒ YES ☐ NO IF ANSWER IS "NO" ADVISE RECOMMENDATION AND/OR ACTION OF SPONSORING OFFICE, ON REVERSE HEREOF. IF KNOWN, INDICATE MAGNITUDE OF ADDITIONAL FUNDS INVOLVED.

HAS AN INTERIM REPORT, FINAL REPORT, PROTOTYPE, OR OTHER END ITEM BEEN RECEIVED FROM THE CONTRACTOR DURING THE PERIOD? ☒ YES ☐ NO (If yes, give details on reverse side.)

HAS GOVERNMENT-OWNED PROPERTY BEEN DELIVERED TO CONTRACTOR DURING THIS PERIOD? ☒ YES ☐ NO (If yes, indicate items, quantity, and cost on reverse side.)

INCENTIVES

IS THIS AN INCENTIVE CONTRACT
IF YES, CHECK TYPE

☐ YES

☒ NO

☐

COST

☐

AWARD
FEE

☐

PERFORMANCE

☐

DELIVERY

NOTE:
USE REVERSE SIDE FOR COMMENTS.
FINAL REPORT MUST CONTAIN INCENTIVE EVALUATION.

OVERALL PERFORMANCE OF CONTRACTOR

1. ☐

OUTSTANDING

4. ☒

ABOVE AVERAGE

7. ☐

UNSATISFACTORY

2. ☐

EXCELLENT

5. ☐

AVERAGE

3. ☐

VERY GOOD

6. ☐

MINIMUM ACCEPTABLE

IF OVERALL PERFORMANCE OF CONTRACTOR IS UNSATISFACTORY OR MINIMUM ACCEPTABLE INDICATE REASONS ON REVERSE SIDE.

RECOMMENDED ACTION

☒

CONTINUE AS PROGRAMMED

☐

WITHHOLD PAYMENT PENDING
SATISFACTORY PERFORMANCE

☐

CLOSE OUT

☐

OTHER (Specify)

IF THIS IS A FINAL REPORT PUT COMMENTS ON REVERSE IN NARRATIVE FORM ON CONTRACTOR'S PERFORMANCE AND CERTIFY THAT ALL DELIVERABLE ITEMS UNDER THE CONTRACT HAVE BEEN RECEIVED. THESE INCLUDE, WHERE APPLICABLE, THE FOLLOWING:

ITEM	REC'D	DOES NOT APPLY	ITEM	REC'D	DOES NOT APPLY
PROTOTYPES			MANUALS		
DRAWINGS AND SPECIFICATIONS			FINAL REPORT		
PRODUCTION AND/OR OTHER END ITEMS			SPECIAL TOOLING		
			OTHER GOVERNMENT PROPERTY		

DATE OF LAST CONTACT WITH CONTRACTOR

27 September 1968

SIGNATURE OF INSPECTOR

DIVISION

TSSG/DED

SIGNATURE OF APPROVER

Approved For Release 2005/05/02 : CIA-RDP78B04770A001200010092-8

FORM
12-67 1897
USE
PREVIOUS
EDITION

☐ UNCLASSIFIED

☐ CONFIDENTIAL

☐ SECRET

(12-36)

☐ SECRET

☐ CONFIDENTIAL

☐ UNCLASSIFIED

Approved For Release 2005/05/02 : CIA-RDP78B04770A001200010092-8

☒ INTERIM

☐ FINAL

- 25X1
1. The Progress Reports for July and August 1968 have been received on time.
- 25X1
2. A ☐ High-Power Stereoviewer and ☐ Eyepieces and Objectives has been furnished to the Contractor as GFI.
- 25X1
3. During an inspection visit of 26-27 September 1968, the following items were discussed:
- 25X1
- a. Because of ☐ failure to place the order for the optical system--partly caused by the time required by the Government to review the specifications--as quickly as initially planned, the contractor plans to request a time extension of the final delivery date but with no increase in cost. In addition, they will request progress payments.
- 25X1
- b. The film stage size will be at least 6 1/8" and probably 6 1/4" (The specifications only require 6").
- 25X1
- c. The human factor data prepared by ☐ was thoroughly discussed. ☐ will attempt to provide the eyepiece tilt from 0° to 23° but in any case will at least provide 8° to 23° range (with 15° midpoint). They were cautioned to avoid image rotation with IPD or eyepiece tilt changes when the optical system was so modified.
- 25X1
- d. ☐ thinks their joystick will be provided with a two step cam: one step providing a logarithmic speed rate for fine positioning and another faster cam for fast, gross movement. The rate control and film stage switch will be located near the joystick.
- 25X1
- e. The knee well will have a minimum depth of 18 inches.
- 25X1
- f. ☐ will investigate the possibility of placing the character readout switches on the comparator itself.
- 25X1
- g. ☐ plans to furnish a chair with the instrument.
- 25X1
- h. ☐ will attempt to provide a writing shelf on the electronics cabinet.

☐ UNCLASSIFIED

☒ CONFIDENTIAL

☐ SECRET